

INSTRUMENT CALIBRATION CERTIFICATE

Cert: 237

Tag ID L12-001	Serial Number 000094515
Function Indicator	Manufacturer LASCAR
Instrument Type Temperature Indicator	Model Number EL-USB-1
Date of Test 01/20/2012	Test Time
As Found Status Passed	Final Status Passed
Testers TRAVIS	Due Date 01/20/2013
	Procedure Number EL-USB-1

Test Equipment Used

Manufacturer	Model Number	Serial Number	Due Date	Cert. Status
HART SCIENTIFIC	1502A	A1B599	05/19/2012	OK

Derick Debolt

Technician

FEB 07 2012

Date

Steve Wick

Supervisor

FEB 07 2012

Date



Lascar Electronics, Inc. certifies that the above equipment has been calibrated using instrumentation and standards that are traceable to the National Institute of Standards and Technology (NIST) through certification documents on file.

This calibration complies with MIL-STD-45662A and ISO 10012-1 and ANSI/ NCSL Z540-1-1994. Test Uncertainty Ratio \geq 4:1 unless otherwise stated.

This certification cannot be reproduced, except in full, without the specific written permission of Lascar Electronics, Inc.



4258 W. 12th Street, Erie, PA 16505 USA

T: (814) 835-0621

F: (814) 838-8141

E-mail: us-sales@lascarelectronics.com

www.lascarelectronics.com

